

Certificate of Compliance

Certificate: 1880227 Master Contract: 162029

Project: 2668222 **Date Issued:** March 21, 2014

Issued to: Fluke Corporation

6920 Seaway Blvd Everett, WA 98203

USA

Attention: Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cíndy Johnson

Issued by: Cindy Johnson, AScT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to

US Standards

Electrical Tester, Models T+, T+ PRO, T+ CAN and T+ PRO CAN, rated 3Vdc (2 x AAA battery operated).

Notes:

- 1. The above models are Equipment Class II, Pollution Degree 2, Measurement Category 1000V CAT III, 600V CAT IV.
- 2. Mode of operation: Continuous for ≤240V; Duty cycle of 30 sec. on 300 seconds off >240V.
- 3. Environmental Conditions: Extended -10 °C to 55 °C, 2000 m max, 0 °C to 30 °C 90%,

30 °C to 40 °C 75%, 40 °C to 50 °C: 45%.

DQD 507 Rev. 2012-05-22 Page: 1



Certificate: 1880227 Master Contract: 162029

Project: 2668222 **Date Issued:** March 21, 2014

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards:

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

DQD 507 Rev. 2012-05-22 Page: 2